



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2015-05
KIT DESCRIPTION:	H-FABP CALIBRATOR SET
CATALOGUE NO:	FB3134
BATCH NO:	353335
EXPIRY DATE:	2017-04

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 098FB	H-FABP	5.81	4.65-6.97	5.54	NG/ML
LEVEL II / 099FB	H-FABP	29.64	23.71-35.57	31.47	NG/ML

APPROVED BY:

Lisa Reer

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S. P. Fitzgerald BSc PhD DSc (Hons) FRAClin. FJ Fitzgerald. W. Fitzgerald DBS. R. P. Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15758, VAT Number GB 353 8301 00.