



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

<b>DATE OF MANUFACTURE:</b>	<b>2016-06</b>
<b>KIT DESCRIPTION:</b>	<b>LIQUID PROTEIN CALIBRATOR</b>
<b>CATALOGUE NO:</b>	<b>IT2691</b>
<b>BATCH NO:</b>	<b>387399</b>
<b>EXPIRY DATE:</b>	<b>2018-01</b>

<b>SAMPLE</b>	<b>TARGET</b>	<b>RANGE</b>	<b>RESULT</b>	<b>UNITS</b>
<b>TRANSFERRIN LEVEL I/ 348LPC</b>	<b>126</b>	<b>101-151</b>	<b>123</b>	<b>mg/dl</b>
<b>LEVEL III/ 350LPC</b>	<b>378</b>	<b>302-454</b>	<b>412</b>	<b>mg/dL</b>

**APPROVED BY:**

**QUALITY CONTROL REPRESENTATIVE**

**Randox Laboratories Limited**, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
**T** +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** [marketing@randox.com](mailto:marketing@randox.com) **I** [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

4706-QC Rev4  
27<sup>th</sup> September 2013