

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-04
KIT DESCRIPTION:	APOLIPOPROTEIN CALIBRATOR
CATALOGUE NO:	LP3023
BATCH NO:	1246LP
EXPIRY DATE:	2018-10-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 1246LP	APO A	241	N/A	240	mg/dl
LEVEL III / 1246LP	APO B	208	N/A	212	mg/dL

APPROVED BY:

Lisa Rea

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S F Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald OBE, R F Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15739, VAT Number GB 353 0904 00.