

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-10
KIT DESCRIPTION:	CYSTATIN C CONTROL LEVEL III
CATALOGUE NO:	CYS5020
BATCH NO:	1212CY
EXPIRY DATE:	2018-05

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL III / 1212CY	3.29	2.96 – 3.62	3.35	mg/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim. BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com