

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-03-28
KIT DESCRIPTION:	C.S.F CONTROL LEVEL 2
CATALOGUE NO:	CF1500
BATCH NO:	113CF
EXPIRY DATE:	2020-03-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
113CF	SODIUM	128.0	122.0-134.0	126	mmol/L
113CF	CHLORIDE	86.7	79.8-93.6	83	mmol/L

APPROVED BY:



QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI5738. VAT Number GB 353 0304 00.