



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-04-28
KIT DESCRIPTION:	COAGULATION CONTROL LEVEL 1
CATALOGUE NO:	CG5021
BATCH NO:	242CG
EXPIRY DATE:	2019-04-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
242CG	PROTHROMBIN TIME	14.4	11.5-17.3	13.9	SECS
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY:

Yvette Miller

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path R J Fitzgerald W Fitzgerald OBE B F Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0301 00.