

CERTIFICATE OF ANALYSIS

This is to certify that the following kits were tested by Randox Testing Services and approved by the QC Department.

DATE OF MANUFACTURE: November 2015

KIT DESCRIPTION: Immunoassay Premium Plus Level 1

CATALOGUE NO: IA3109

LOT NUMBERS: 1506EC **EXPIRY DATE:** 2018-04

SYSTEM OF ANALYSIS: Siemens Immulite 2000Xpi

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FSH (U/L)	1506EC	5.27 - 7.91	6.59	7.0	
				6.7	
LH (U/L)	1506EC	4.19 - 6.29	5.24	5.7	
				6.0	
TSH (mIU/I)	1506EC	0.102 - 0.154	0.128	0.128	
				0.135	

QC TESTED BY:		Aimee Anderson		
DATE:	24-Nov-15			
RESULT (PASS/FAIL):	Pass		
QC APPROVED BY:		adèle Martin.		

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com