

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: July 2016
KIT DESCRIPTION: IA PREMIUM PLUS 2
CATALOGUE NO: IA3110
LOT NUMBERS: 1621EC
EXPIRY DATE: 2019-04-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000 xpi

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Estradiol pmol/l	1621EC	825-1237	1031	1083	
				1061	
FSH U/l	1621EC	21.1-31.7	26.4	27.9	
				27.3	
Prolactin mIU/l	1621EC	554-830	692	727	
				736	

QC TESTED BY: Aimee Anderson
DATE: 29 July 2016
RESULT (PASS/FAIL): Pass
QC APPROVED BY: *Adele Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15736, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016