

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: April 2016
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS Level 2
CATALOGUE NO: IA3110
LOT NUMBERS: 1644EC
EXPIRY DATE: 2019-04-18
SYSTEM OF ANALYSIS: Siemens Immulite 2000 Xpi

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Cortisol (nmol/l)	1644EC	441 - 735	588	612	N/A
				632	
SHBG (nmol/l)	1644EC	31.4 - 47.2	39.3	38.5	N/A
				38.6	
FT3 (pmol/l)	1644EC	8.10 - 13.5	10.8	11.4	N/A
				11.0	

QC TESTED BY: Davina Graham
DATE: 12 Jan 2017
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *Adele Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE Man. R. J. Fitzgerald, W. Fitzgerald OBE, R. F. Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15736. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016