

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: April 2016  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS 3  
CATALOGUE NO: IA3111  
LOT NUMBERS: 1646EC  
EXPIRY DATE: 2019-04-28  
SYSTEM OF ANALYSIS: Siemens Immulite 2000 Xpi

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (KIU/l)	1646EC	157 - 235	196	188	
				181	
SHBG (nmol/l)	1646EC	32.1 - 48.1	40.1	39.6	
				41.6	
TSH (mIU/l)	1646EC	17.0 - 25.6	21.3	20.4	
				21.0	

QC TESTED BY: Davina Graham  
DATE: 17 Jan 2017  
RESULT (PASS/FAIL): PASS  
QC APPROVED BY: *Adele Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15736, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016