

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2017-05-28  
KIT DESCRIPTION: Maternal Control Level 3  
CATALOGUE NO: MSS5026  
LOT NUMBERS: 6759MS  
EXPIRY DATE: 2020-05-28  
SYSTEM OF ANALYSIS: Siemens Immulite 2000

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (kU/l)	6759MS	77.5 – 116.0	96.9	87.0	N/A
				89.9	
Free Beta hCG (U/L)	6759MS	80.3 – 134.0	107	116	
				105	

QC TESTED BY: Ilona Palmer  
DATE: 29 Nov 2017  
RESULT (PASS/FAIL): Pass  
QC APPROVED BY: *Adele Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) RAe Man, R J Fitzgerald, W Fitzgerald OBE, P P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016