RANDOX

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-04
KIT DESCRIPTION:	CRP CALIBRATOR SERIES
CATALOGUE NO:	CP2499
BATCH NO:	413359
EXPIRY DATE:	2018-09-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 401LPC	FRCRP	24.0	19.2-28.8	22.9	mg/L
LEVEL II / 402LPC	FRCRP	48.4	38.7-58.1	44.5	mg/L

APPROVED BY:

Lisa Rea

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Mem. R J Fitzgerald W Fitzgerald OBE R F Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00

FORM No. 9323-QC
REVISION (1)
27th September 2013