


RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-05-28
KIT DESCRIPTION:	Rheumatoid Factor Standard (RF CAL)
CATALOGUE NO:	RF2301
BATCH NO:	441079
EXPIRY DATE:	2018-11-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL I 488LPC	28.8	23.0 – 34.6	29.6	U/ml
LEVEL II 463LPC	39.2	31.4 – 47.0	39.4	U/ml

APPROVED BY:



QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald CBE, P P Kelly BSc FRC Path
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 8304 00