

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: April 2017
KIT DESCRIPTION: IA Premium Plus 3
CATALOGUE NO: IA3111
LOT NUMBERS: 1719EC
EXPIRY DATE: 2020-04-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Prolactin (uU/ml)	1719EC	1138-1708	1423	1406	N/A
				1389	
FSH (U/L)	1719EC	49.0-73.4	61.2	59.7	N/A
				59.4	
CEA (ug/l)	1719EC	65.9-98.9	82.4	85.9	N/A
				81.8	

QC TESTED BY: Ilona Palmer
DATE: 12 Oct 2017
RESULT (PASS/FAIL): Pass

QC APPROVED BY:

Adele Martin

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MSc R J Fitzgerald, W Fitzgerald OBE R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15736, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016