

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a  
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: April 2017  
KIT DESCRIPTION: IA Premium Plus 2  
CATALOGUE NO: IA3110  
LOT NUMBERS: 1727EC  
EXPIRY DATE: 2020-04-28  
SYSTEM OF ANALYSIS: Siemens Immulite 2000

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Prolactin (uU/ml)	1727EC	526-790	658	685	N/A
				676	
FSH (U/L)	1727EC	21.1-31.7	26.4	28.1	N/A
				26.2	
CEA (ug/l)	1727EC	30.2-45.2	37.7	39.8	N/A
				36.7	

QC TESTED BY: Ilona Palmer  
DATE: 12 Oct 2017  
RESULT (PASS/FAIL): Pass  
QC APPROVED BY: Adèle Martin.

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE Man. R J Fitzgerald, W Fitzgerald OBE R F Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016