



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-09-29
KIT DESCRIPTION:	Liquid Protein Calibrators
CATALOGUE NO:	IT2691
BATCH NO:	466110
EXPIRY DATE:	2019-09-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level II/ 496LPC	Transferrin	263	210 – 316	262	mg/dL
Level III/ 510LPC	Transferrin	391	313 – 469	383	mg/dL

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRCM, R J Fitzgerald, W Fitzgerald OBE, R P Kealy BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0904 00.