



## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-10-28
KIT DESCRIPTION:	APOLIPOPROTEIN CALIBRATOR
CATALOGUE NO:	LP3023
BATCH NO:	1348LP
EXPIRY DATE:	2019-10-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
1348LP	APO A	241	N/A	244	mg/dL
1348LP	APO B	208	N/A	208	mg/dL

APPROVED BY:

*Yvette Miller*

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MRC R J Fitzgerald W Fitzgerald OBE R F Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.