



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

| | |
|-----------------------------|--------------------------------|
| DATE OF MANUFACTURE: | 2018-03-29 |
| KIT DESCRIPTION: | hsCRP CALIBRATOR SERIES |
| CATALOGUE NO: | CP2478 |
| BATCH NO: | 460831 |
| EXPIRY DATE: | 2020-03-28 |

| SAMPLE | ANALYTE | TARGET | RANGE | RESULT | UNITS |
|---|----------------|---------------|------------------|---------------|--------------|
| hsCRP CONTROL LEVEL II/ 2528 CP | hsCRP | 1.02 | 0.77-1.28 | 1.01 | mg/L |
| hsCRP CONTROL LEVEL III/ 2430 CP | hsCRP | 4.34 | 3.47-5.21 | 4.16 | mg/L |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Mar. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.