



## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-09-29
KIT DESCRIPTION:	CRP CONTROL LEVEL 2
CATALOGUE NO:	CP2480
BATCH NO:	2599CP
EXPIRY DATE:	2020-03-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
CRP CONTROL LEVEL II / 2599CP	CRP	19.46	N/A	20.59	mg/L
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) | [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MSc, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NH5738, VAT Number GB 353 6304 00.