



HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-07-29
KIT DESCRIPTION:	TRI- LEVEL CARDIAC CONTROL
CATALOGUE NO:	CQ3259
BATCH / LOT NO:	432213
EXPIRY DATE:	2021-07-28

ANALYTE	TARGET	RANGE	RESULT	UNITS
CKMB / 4173CK	20.6	16.5-24.7	21.0	U/L
CKMB / 4174CK	136.0	109.0-163.0	130.6	U/L
CKNAC / 4172CK	65.0	53.0-77.0	70.7	U/L
CKNAC / 4173CK	233.0	191.0-275.0	220.7	U/L
CKNAC / 4174CK	600.0	492.0-708.0	593.3	U/L

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2412 F +44 (0) 28 9445 2912 E marketing@randox.com | www.randox.com

Traders: Dr S P Fitzgerald BSc PhD DSc (Hons) FRACRian D] Fitzgerald, W Fitzgerald QMB DPHist BSc FRCR
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI 13798. VAT Number GB 353 0524 00

4705-QC RevB
27th September 2013