

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-05-29
KIT DESCRIPTION: IA Premium Plus Level II
CATALOGUE NO: IA3110
LOT NUMBERS: 1812EC
EXPIRY DATE: 2021-05-28
SYSTEM OF ANALYSIS: Roche Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Alphafoetoprotein kU/l	1812EC	39.8-59.6	49.7	50.5	N/A
				50.0	
CEA (ug/l)	1812EC	17.2-25.8	21.5	21.0	N/A
				20.9	
Total HCG+B mU/ml	1812EC	12.0-18.0	15.0	15.1	N/A
				15.2	

QC TESTED BY: Davina Graham
DATE: 21-Nov-18
RESULT (PASS/FAIL): Pass
QC APPROVED BY: *C. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc PCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016