

LIQUID CARDIAC CONTROL - LEVEL 2 (CRD LIQ CONTROL 2)

Cat. No. CQ5052 Lot No. 4312CK Size: 3 x 3 ml Expiry: 2020-06-28

Range					
Analyte	unit	Target	low	high	methods
CK-MB Mass 肌酸激酶同工酶 (质量)	ng/ml = µg/l	16.5	11.6	21.5	Abbott Architect
	ng/ml = µg/l	22.9	16.0	29.8	Siemens Centaur XP/XPT/Classic
	ng/ml = µg/l	22.3	15.6	29.0	Siemens Dimension
	ng/ml = µg/l	15.1	10.6	19.6	Roche Elecsys Modular E170 Cobas 6000/e411
	ng/ml = µg/l	23.9	16.7	31.1	Beckman Coulter Access
	ng/ml = µg/l	21.5	15.1	28.0	Siemens Stratus CS
	ng/ml = µg/l	21.2	14.8	27.6	BioMerieux Vidas
	ng/ml = µg/l	24.0	16.8	31.2	Beckman DxI800
	ng/ml = µg/l	14.6	10.2	19.0	Roche h232
	ng/ml = µg/l	30.0	21.0	39.0	Radiometer AQT90 Flex
	ng/ml = µg/l	20.1	14.1	26.1	Siemens Centaur CP
D-Dimer D-二聚体	µg/l FEU	1384	1038	1730	Biomerieux Vidas Exclusion II
	µg/l FEU	5244	3933	6555	Mitsubishi Pathfast D-Dimer
	µg/l	539	404	674	Roche/ Stago STA-R Evolution
	µg/l	829	622	1036	Roche Cobas h232 D-Dimer
	µg/l	459	344	574	Roche Integra D-DI 2
	µg/l	712	534	890	Alere Biosite Triage D-Dimer
	µg/l	726	545	908	Abbott Architect Quantia D-Dimer
	µg/l	848	636	1060	Siemens Stratus CS
	µg/l	763	572	954	Radiometer AQT90 Flex D-Dimer
	µg/l FEU	1868	1401	2335	Siemens Innovance D-Dimer
	µg/l	351	263	439	Roche Cobas D-DI 2
	µg/l FEU	2122	1592	2653	HemosIL D-Dimer HS 500
	µg/l	549	412	686	HemosIL D-Dimer
Digoxin 地高辛	nmol/l	2.28	1.82	2.74	Chemiluminescence
	ng/ml	1.78	1.42	2.14	
	nmol/l	2.19	1.75	2.63	Enzyme Immunoassay
	ng/ml	1.71	1.37	2.05	
	nmol/l	2.28	1.82	2.74	Turbidimetric
	ng/ml	1.78	1.42	2.14	
	nmol/l	2.17	1.74	2.60	KIMS
	ng/ml	1.69	1.36	2.02	
hsCRP 高敏C反应蛋白	mg/l	2.85	2.28	3.42	Nephelometric (IFCC Cal.)
	mg/l	2.95	2.36	3.54	Turbidimetric (IFCC Cal.)
	mg/l	2.99	2.39	3.59	Turbidimetric (Non IFCC Cal.)
	mg/l	2.76	2.21	3.31	Randox Immunoturbidimetric
Myoglobin 肌红蛋白	ng/ml = µg/l	140	98.0	182	Abbott Architect
	ng/ml = µg/l	104	72.8	135	Siemens/Dade Behring Nephelometer
	ng/ml = µg/l	138	96.6	179	Siemens Centaur XP/XPT/Classic
	ng/ml = µg/l	139	97.3	181	Siemens Dimension
	ng/ml = µg/l	95.6	66.9	124	Beckman DxI800
	ng/ml = µg/l	105	73.5	137	Roche Elecsys

LIQUID CARDIAC CONTROL - LEVEL 2 (CRD LIQ CONTROL 2)

Cat. No. CQ5052 Lot No. 4312CK Size: 3 x 3 ml Expiry: 2020-06-28

Range					
Analyte	unit	Target	low	high	methods
Myoglobin 肌红蛋白	ng/ml = µg/l	106	74.2	138	Roche Hitachi
	ng/ml = µg/l	93.7	65.6	122	Beckman Coulter Access
	ng/ml = µg/l	71.5	50.1	93.0	Siemens Stratus CS
	ng/ml = µg/l	108	75.6	140	BioMerieux Vidas
	ng/ml = µg/l	111	77.7	144	Siemens Dimension Vista LOCI
	ng/ml = µg/l	123	86.1	160	Siemens Centaur CP
	ng/ml = µg/l	155	109	202	Randox Immunoturbidimetric
NT-ProBNP N端脑钠肽前体	pmol/l	40.6	30.5	50.8	Siemens Centaur XP/XPT/Classic
	pg/ml	344	258	430	
	pmol/l	216	162	270	Siemens Immulite 2000
	pg/ml	1830	1372	2288	
	pmol/l	55.2	41.4	69.0	Siemens Stratus CS
	pg/ml	468	351	585	
	pmol/l	59.5	44.6	74.4	BioMerieux Vidas
	pg/ml	504	378	630	
	pmol/l	45.6	34.2	57.0	Roche Elecsys Modular E170 Cobas 6000/e411
	pg/ml	386	290	482	
	pmol/l	159	119	199	Mitsubishi Chemical Pathfast
	pg/ml	1347	1008	1686	
	pmol/l	93.9	70.4	117	Ortho Vitros 3600/5600/ECi
	pg/ml	796	596	996	
	pmol/l	27.1	20.3	33.9	Roche h232
	pg/ml	230	172	288	
	pmol/l	18.3	13.7	22.9	Siemens Dimension Vista LOCI
	pg/ml	155	116	194	
	pmol/l	9.80	7.35	12.3	Siemens Dimension Exl LOCI
pg/ml	83.0	62.3	104		
pmol/l	59.9	44.9	74.9	Biomerieux Vidas 2	
pg/ml	507	380	634		
Troponin I 肌钙蛋白I	ng/ml = µg/l	2.02	1.62	2.42	Siemens Centaur XP/XPT/Classic
	ng/l = pg/ml	2020	1620	2420	
	ng/ml = µg/l	1.10	0.880	1.32	Siemens Dimension
	ng/l = pg/ml	1100	880	1320	
	ng/ml = µg/l	1.21	0.968	1.45	Siemens Stratus CS
	ng/l = pg/ml	1210	968	1452	
	ng/ml = µg/l	8.76	7.01	10.5	Ortho Vitros ECi
	ng/l = pg/ml	8760	7010	10510	
	ng/ml = µg/l	7.87	6.30	9.44	Biomerieux Vidas Ultra
	ng/l = pg/ml	7870	6300	9440	
	ng/ml = µg/l	0.603	0.482	0.724	Roche Elecsys/E170/c6000/e411
	ng/l = pg/ml	603	482	724	
	ng/ml = µg/l	0.721	0.577	0.865	Mitsubishi Chemical Pathfast
	ng/l = pg/ml	721	577	865	
	ng/ml = µg/l	0.949	0.759	1.14	Siemens/Dade Dimension EXL/Vista
ng/l = pg/ml	949	759	1139		
ng/ml = µg/l	1.04	0.832	1.25	Siemens Dimension Exl LOCI	
ng/l = pg/ml	1040	832	1248		

LIQUID CARDIAC CONTROL - LEVEL 2 (CRD LIQ CONTROL 2)

Cat. No. CQ5052 Lot No. 4312CK Size: 3 x 3 ml Expiry: 2020-06-28

Range					
Analyte	unit	Target	low	high	methods
Troponin I 肌钙蛋白I	ng/ml = µg/l	1.65	1.32	1.98	Abbott Architect STAT hs
	ng/l = pg/ml	1650	1320	1980	
	ng/ml = µg/l	0.958	0.766	1.15	Beckman Dxl - AccuTnl+3
	ng/l = pg/ml	958	766	1150	
	ng/ml = µg/l	1.03	0.824	1.24	Beckman Access - AccuTnl+3
	ng/l = pg/ml	1030	824	1236	
	ng/ml = µg/l	1.59	1.27	1.91	Siemens Centaur CP
	ng/l = pg/ml	1590	1270	1910	
	ng/ml = µg/l	7.84	6.27	9.41	bioMerieux VIDAS hs Troponin I
	ng/l = pg/ml	7840	6270	9410	
ng/ml = µg/l	3.26	2.61	3.91	Siemens Centaur XP/XPT High Sensitivity Troponin I (TNIH)	
ng/l = pg/ml	3260	2610	3910		