



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-08-29
KIT DESCRIPTION:	TRI-LEVEL CARDIAC CONTROL
CATALOGUE NO:	CQ3259
BATCH NO:	483221
EXPIRY DATE:	2022-08-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL I/ 4324CK	CK-NAC	78	64-92	73	U/L
LEVEL II/ 4325CK	CK-NAC	257	211-303	241	U/L
LEVEL III/ 4326CK	CK-NAC	657	539-775	613	U/L
LEVEL II/ 4325CK	CK-MB	21.4	17.1-25.7	19.2	U/L
LEVEL III/ 4326CK	CK-MB	123.0	98.4-148.0	114.5	U/L

APPROVED BY:

Lisa Murray

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.