

## RANDEX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2018-03-29
KIT DESCRIPTION:	MICROALBUMIN (NALB)
CATALOGUE NO:	MA1361
BATCH NO:	448770
EXPIRY DATE:	2020-03-29

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL 1 S4486A	MICROALBUMIN	32.1	25.7 - 38.5	30.4	MG/L
LEVEL 2 S4487A	MICROALBUMIN	161.2	129.0 - 193.5	158.9	MG/L

APPROVED BY:



QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 95 Clonmel Road, Crumlin, County Dubain, BT27 4QJ United Kingdom  
T +44 (0) 28 9442 2000 F +44 (0) 28 9442 2002 E [marketing@randox.co.uk](mailto:marketing@randox.co.uk) I [www.randox.com](http://www.randox.com)

Developed and approved by RANDEX using the real-time process management system. For more information, contact Randox Laboratories Limited. All reagents supplied to the user. Randox Certificates are issued for the 100% of the product.