

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-05-29
KIT DESCRIPTION: IA SPECIALITY II LEV 2
CATALOGUE NO: IAS3118
LOT NUMBERS: 1800EC
EXPIRY DATE: 2020-05-28
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (ng/l)	1800EC	84.6 - 141	113	110	N/A
				109	
Gastrin (pmol/l)	1800EC	34.7 - 57.8	46.2	48.1	N/A
				48.1	

QC TESTED BY: Davina Graham
DATE: 19-Sep-19
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *P. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number: NI15738. VAT Number: GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016