



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

<b>DATE OF MANUFACTURE:</b>	<b>2018-11-29</b>
<b>KIT DESCRIPTION:</b>	<b>H-FABP CALIBRATOR</b>
<b>CATALOGUE NO:</b>	<b>FB3134</b>
<b>BATCH NO:</b>	<b>472395</b>
<b>EXPIRY DATE:</b>	<b>2020-04-28</b>

<b>SAMPLE</b>	<b>ANALYTE</b>	<b>TARGET</b>	<b>RANGE</b>	<b>RESULT</b>	<b>UNITS</b>
<b>LEVEL II 152FB</b>	<b>H-FABP</b>	<b>7.33</b>	<b>5.86 – 8.80</b>	<b>7.27</b>	<b>ng/ml</b>
<b>LEVEL III 153FB</b>	<b>H-FABP</b>	<b>30.41</b>	<b>24.33 – 36.49</b>	<b>31.93</b>	<b>ng/ml</b>

APPROVED BY:

**QUALITY CONTROL REPRESENTATIVE**

**Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom**  
**T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)**

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R F Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.