

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a  
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-10-29  
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY 2 - LEVEL 1  
CATALOGUE NO: IAS3117  
LOT NUMBERS: 1854EC  
EXPIRY DATE: 2020-10-28  
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (pg/ml)	1854EC	18.4-30.6	24.5	26.0	N/A
				24.2	
Gastrin (pmol/l)	1854EC	19.7-32.8	26.2	28.5	N/A
				29.0	

QC TESTED BY: Davina Graham  
DATE: 03-Dec-19  
RESULT (PASS/FAIL): PASS  
QC APPROVED BY: *B. McAteer*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016