

## **CERTIFICATE OF ANALYSIS**

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE:

2018-10-29

KIT DESCRIPTION:

IMMUNOASSAY SPECIALITY II - LEVEL 2

**CATALOGUE NO:** 

IAS3118

LOT NUMBERS:

1855EC

**EXPIRY DATE:** 

2020-10-28

SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

## **TEST CONTROLS OR VIALS**

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
	1855EC	93.1 - 155	124	122	N/A
Calcitonin (pg/ml)				125	
	1855EC	42.2 - 70.4	56.3	61.0	N/A
Gastrin (pmol/l)				64.8	

QC TESTED BY:		Davina Graham		
DATE:	12-Dec-19			
RESULT	(PASS/FAIL):	PASS		
QC APPROVED BY:		a. Manten		

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com