

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-04-29  
KIT DESCRIPTION: Maternal Control 3  
CATALOGUE NO: MSS5026  
LOT NUMBERS: 7140MS  
EXPIRY DATE: 2021-04-28  
SYSTEM OF ANALYSIS: Roche Cobas e801

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Alphafoetoprotein kU/l	7140MS	100-150	125.0	122.0	N/A
				120.0	
Total Beta hCG (mU/ml)	7140MS	22011-33017	27514	29433	N/A
				30163	

QC TESTED BY: Aimee Anderson  
DATE: 06-Dec-18  
RESULT (PASS/FAIL): Pass  
QC APPROVED BY: *E. Menten*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016