

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-04-29
KIT DESCRIPTION: MATERNAL CONTROL 3
CATALOGUE NO: MSS5026
LOT NUMBERS: 7462MS
EXPIRY DATE: 2022-04-28
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (IU/ml)	7462MS	88.0 - 132	110	118.0	N/A
				118.7	
Total Beta hCG (U/l)	7462MS	43217-64825	54021	54366	N/A
				52776	

QC TESTED BY: Davina Graham
DATE: 13-Sept-19
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016