

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-05-29  
KIT DESCRIPTION: IA PREMIUM PLUS LEVEL 3  
CATALOGUE NO: IA3111  
LOT NUMBERS: 1815EC  
EXPIRY DATE: 2021-05-28  
SYSTEM OF ANALYSIS: Roche Cobas e801

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Alphafoetoprotein (kU/l)	1815EC	138 - 208	173	175	N/A
				177	
CEA (ug/l)	1815EC	38.6 - 57.8	48.2	49.5	N/A
				49.2	
total PSA (ug/l)	1815EC	23.1 - 38.5	30.8	30.8	N/A
				31.3	

QC TESTED BY: Davina Graham  
DATE: 14-Nov-18  
RESULT (PASS/FAIL): Pass  
QC APPROVED BY: *R. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016