

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-05-29
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 3
CATALOGUE NO: IA3111
LOT NUMBERS: 1913EC
EXPIRY DATE: 2022-05-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Cortisol (nmol/l)	1913EC	626 - 1044	835	720	N/A
				745	
DHEA-S (umol/l)	1913EC	15.0 - 22.4	18.7	16.2	N/A
				16.8	
FSH (mU/m/l)	1913EC	48.4 - 72.6	60.5	61.0	N/A
				60.5	

QC TESTED BY: Davina Graham
DATE: 10-Dec-19
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *G. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016