

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-10-29

KIT DESCRIPTION: IMMUNOASSAY SPECIALITY II LEVEL 1

CATALOGUE NO: IA3117

LOT NUMBERS: 1949EC

EXPIRY DATE: 2021-10-28

SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

RESULTS FOR TEST CONTROLS OR VIALS	LOT NO	RANGE	TARGET	RESULT	PRECISION (if applicable)
ANALYTE					
Calcitonin (ng/l)	1949EC	21.0-35.2	28.1	29.1	
				28.4	
Gastrin (pmol/l)	1949EC	12.2-20.3	16.2	15.9	
				16.0	
PASS or FAIL	PASS				

QC TESTED BY: Emer Ni Bhraonain

DATE: 2-June-20

RESULT (PASS/FAIL): Pass

QC APPROVED BY: 

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE. R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number: NI15738. VAT Number: GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016