

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-10-29
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY 2 LEVEL 3
CATALOGUE NO: IAS3119
LOT NUMBERS: 1951EC
EXPIRY DATE: 2021-10-28
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

| RESULTS FOR TEST CONTROLS OR VIALS | LOT NO | RANGE | TARGET | RESULT | PRECISION (if applicable) |
|------------------------------------|--------|---------|--------|--------|---------------------------|
| ANALYTE | | | | | |
| CALCITONIN (pg/ml) | 1951EC | 250-418 | 334 | 336 | |
| | | | | 337 | |
| GASTRIN (pg/ml) | 1951EC | 174-288 | 231 | 237 | |
| | | | | 231 | |

QC TESTED BY: Emer Ni Bhraonain
DATE: 20-Apr-20
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *[Signature]*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA,
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016