



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

<b>DATE OF MANUFACTURE:</b>	<b>2019-08-29</b>
<b>KIT DESCRIPTION:</b>	<b>H-FABP CONTROL LEVEL 1</b>
<b>CATALOGUE NO:</b>	<b>FB4026</b>
<b>BATCH NO:</b>	<b>192FB</b>
<b>EXPIRY DATE:</b>	<b>2021-01-28</b>

<b>SAMPLE</b>	<b>ANALYTE</b>	<b>TARGET</b>	<b>RANGE</b>	<b>RESULT</b>	<b>UNITS</b>
<b>LEVEL I / 192FB</b>	<b>H-FABP</b>	<b>7.64</b>	<b>6.11-9.17</b>	<b>7.28</b>	<b>ng/ml</b>

**APPROVED BY:**

*Lisa Murray*

**QUALITY CONTROL REPRESENTATIVE**

**Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) | [www.randox.com](http://www.randox.com)**

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MInst R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.