

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-04-29
KIT DESCRIPTION: IA Speciality 1 Lev 2
CATALOGUE NO: IAS3114
LOT NUMBERS: 1796EC
EXPIRY DATE: 2020-04-28
SYSTEM OF ANALYSIS: Roche Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Thyroglobulin antibodies (U/ml)	1796EC	662-1103	882	956	N/A
				897	
Insulin (mU/l)	1796EC	14.2-23.6	18.9	19.4	N/A
				20.0	
C-Peptide (ng/ml)	1796EC	3.53-5.89	4.71	4.60	N/A
				4.71	

QC TESTED BY: Tara Smith
DATE: 12-Dec-18
RESULT (PASS/FAIL): Pass
QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016