

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-05-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 2  
CATALOGUE NO: IA3110  
LOT NUMBERS: 1910EC  
EXPIRY DATE: 2022-05-28  
SYSTEM OF ANALYSIS: Siemens Immulite 2000

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Cortisol (nmol/l)	1910EC	412 - 686	549	488	N/A
				516	
Thyroglobulin (ng/ml)	1910EC	12.1 - 20.1	16.1	17.8	N/A
				17.9	
Total T4 (nmol/l)	1910EC	107 - 178	142	133	N/A
				126	

QC TESTED BY: Davina Graham  
DATE: 21-Nov-19  
RESULT (PASS/FAIL): PASS  
QC APPROVED BY: *G. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S.P. Fitzgerald BSc PhD DSc (Hons) FIAE Man. R.J. Fitzgerald, W. Fitzgerald OBE, R.P. Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016