

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-11-29
KIT DESCRIPTION: IA PREMIUM PLUS 1
CATALOGUE NO: IA3109
LOT NUMBERS: 1862EC
EXPIRY DATE: 2021-11-28
SYSTEM OF ANALYSIS: Roche Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Alphafetoprotein (KU/l)	1862EC	8.8 - 13.2	11.0	10.7	N/A
				10.8	
PSA free (ug/l)	1862EC	0.722 - 1.20	0.962	0.972	N/A
				0.969	
PSA total (ug/l)	1862EC	1.35 - 2.25	1.80	1.68	N/A
				1.70	

QC TESTED BY: Davina Graham
DATE: 5-Jun-19
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016