

RANDOX


CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-11-29
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY I - LEVEL 2
CATALOGUE NO: IAS3114
LOT NUMBERS: 1944EC
EXPIRY DATE: 2021-11-28
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
IGF-1 (µg/l)	1944EC	207-345	276	258	
				260	
Parathyroid Hormone (pmol/l)	1944EC	34.5-57.5	46.0	46.1	
				45.3	

QC TESTED BY: Emer Ni Bhraonain
DATE: 5-Mar-20
RESULT (PASS/FAIL): Pass
QC APPROVED BY: 

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016