



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2020-10-29
KIT DESCRIPTION:	RHEUMATOID FACTOR (RF CAL)
CATALOGUE NO:	RF2301
BATCH NO:	542073
EXPIRY DATE:	2022-04-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II/ 517LPC	58.3	46.6-70.0	64.2	IU/ML
LEVEL III/ 566LPC	109.0	87.2-131.0	112.8	IU/ML

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) | [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald W Fitzgerald CBE. R P Kelly BSc PCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738. VAT Number GB 353 0204 00.