



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

| | |
|-----------------------------|--|
| DATE OF MANUFACTURE: | 2018-12-29 |
| KIT DESCRIPTION: | APOLIPOPROTEIN CALIBRATOR (APO CAL) |
| CATALOGUE NO: | LP3023 |
| BATCH NO: | 1558LP |
| EXPIRY DATE: | 2021-12-28 |

| SAMPLE | ANALYTE | TARGET | RANGE | RESULT | UNITS |
|---------------|----------------|---------------|--------------|---------------|--------------|
| 1558LP | APO A-I | 253 | N/A | 252 | mg/dl |
| 1558LP | APO B | 247 | N/A | 245 | mg/dl |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

**Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | www.randox.com**

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R F Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.