



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2020-02-29
KIT DESCRIPTION:	Blood Gas Control Level 3
CATALOGUE NO:	BG5003
BATCH NO:	214BG
EXPIRY DATE:	2022-02-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level III / 214BG	Direct ISE Sodium	160.00	152.00-168.00	161.27	mmol/L
Level III / 214BG	Direct ISE Potassium	6.11	5.62-6.60	6.06	mmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

**Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com**

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MInt R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NIH 5738, VAT Number GB 353 0304 00.