

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-12-29
KIT DESCRIPTION: Immunoassay Premium Plus Level 3
CATALOGUE NO: IA3111
LOT NUMBERS: 1942EC
EXPIRY DATE: 2022-12-28
SYSTEM OF ANALYSIS: Immulite 2000 xpi

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FSH(U/L)	1942EC	48.4 – 72.6	60.5	57.3	N/A
				60.1	
PRL(mIU/L)	1942EC	758- 1138	948.0	1088	N/A
				1077	
FT3(pmol/l)	1942EC	12.2 – 20.3	16.2	16.6	N/A
				16.3	

QC TESTED BY: EMER NI BHRAONAIN
DATE: 30-Jul-20
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *A. Mculan*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738. VAT Number G8 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016