

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Radox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2020-04-29
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY II - LEVEL 1
CATALOGUE NO: IAS3117
LOT NUMBERS: 1992EC
EXPIRY DATE: 2022-04-28
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Gastrin (pmol/l)	1992EC	17.5 - 29.1	23.3	20.9	N/A
				20.5	
Calcitonin (ng/l)	1992EC	21.1 - 35.1	28.1	28.5	N/A
				28.3	

QC TESTED BY: Emer Ni Bhraonain
DATE: 27-Nov-20
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *B. Meara*

Radox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@radox.com I www.radox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Radox Laboratories Limited is a company registered in Northern Ireland, Company Number NII 5738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016