

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFAC	TURE:	2020-10-29		
KIT DESCRIPTION:	IMMUNO	ASSAY SPECIALITY II - LEVEL 3		
CATALOGUE NO:	IAS 3119			
LOT NUMBERS:	2019EC			
EXPIRY DATE:	2022-10-28			
SYSTEM OF ANALY	SIS: SIEMEN	IS IMMULITE 2000		

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (pg/ml)	2019EC	247-411	329	339	N/A
				336	
(F9)					
Gastrin (pmol/l)	2019EC	131-219	175	183	N/A
				178	
(66)					
	<u> </u>				
		The second secon			

QC TESTED BY:		Davina Graham		
DATE:	28-JAN-21			
RESULT	(PASS/FAIL):	PASS		
QC APP	ROVED BY:	a-Mantey		

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr. 5 P. Fitzgerald BSc. PhD DSc. (Hons) FIAE Man. R. J. Fitzgerald, W. Fitzgerald QRE. R. P. Kelly BSc. FCA, Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NII 5733, VAT. Number. GB 353-0304-00.