

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2020-10-29
KIT DESCRIPTION:	CRP Calibrator (CRP CAL)
CATALOGUE NO:	CP2179
BATCH NO:	2782CP
EXPIRY DATE:	2022-04-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II/ 517LPC	CRP	44.9	35.9-53.9	45.7	mg/L
LEVEL III/ 566LPC	CRP	68.9	55.1-82.7	73.9	mg/L

APPROVED BY:

hisa Murray

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com