



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2020-02-29
KIT DESCRIPTION:	High Sensitivity CRP Calibrator Series
CATALOGUE NO:	CP2478
BATCH NO:	541746
EXPIRY DATE:	2022-02-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL I/ 2728CP	HSCRIP	1.03	0.77-1.29	1.13	mg/L
LEVEL II/ 2611CP	HSCRIP	4.33	3.46-5.20	4.39	mg/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

**Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com**

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI115738, VAT Number GB 353 0304 00.