



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2020-09-29
KIT DESCRIPTION:	Microalbumin Calibrator Series
CATALOGUE NO:	MA1567
BATCH NO:	549476
EXPIRY DATE:	2022-09-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL I/ 1036MA	mALB	33.2	26.6-39.8	34.1	mg/L
LEVEL II/ 1037MA	mALB	165.3	132.2-198.4	177.1	mg/L

APPROVED BY:

Lisa Murray

QUALITY CONTROL REPRESENTATIVE

**Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com**

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE MRB R J Fitzgerald W Fitzgerald CBE R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI115738, VAT Number GB 353 0904 00.