

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2020-11-29
 KIT DESCRIPTION: IMMUNOASSAY SPECIALITY 1 LEVEL 2
 CATALOGUE NO: IAS3114
 LOT NUMBERS: 2013EC
 EXPIRY DATE: 2022-11-28
 SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
IGF-1 (ug/l)	2013EC	187 - 311	249	249	N/A
				238	
Anti TPO (kU/l)	2013EC	96.0 - 160	128	119	N/A
				112	

QC TESTED BY: Charlene Junkin
 DATE: 26-MAR-21
 RESULT (PASS/FAIL): PASS
 QC APPROVED BY: *R. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
 T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
 Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.