

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2019-12-29
KIT DESCRIPTION: Immunoassay Premium Plus Level 3
CATALOGUE NO: IA3111
LOT NUMBERS: 1972EC
EXPIRY DATE: 2022-12-28
SYSTEM OF ANALYSIS: ROCHE COBAS 8000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Total PSA (ug/l)	1972EC	27.3 - 45.5	36.4	35.3	N/A
				35.7	
Insulin (pmol/l)	1972EC	25.4 - 42.3	33.8	32.4	N/A
				32.1	
Free PSA (ug/l)	1972EC	12.9 - 21.5	17.2	17.5	N/A
				17.3	

QC TESTED BY: Aimee Anderson
DATE: 25-Feb-21
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *G. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAC Man R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016